

AS/NZS 4308:2008 - Certificate of Compliance

**DrugCheck Drug Screen Cup – Manufactured by Express Diagnostics
International Inc.
Part No: 60601-6**

The DrugCheck Drug Screen Cup was tested in accordance with Appendix B – “Verification of performance of on-site devices around the cut-off”, of the Australian and New Zealand Standard AS/NZS 4308:2008.

The DrugCheck Drug Screen Cup was shown to satisfy the specified performance criteria when results are interpreted as described in the package inserts (Numbered: DC112F-IVD/August 2013, DC239A-IVD/February 2012, DC242A-IVD/February 2012, DC241A-IVD/March 2012).

Results of the verification procedure used to assess compliance of the DrugCheck Drug Screen Cup have been documented and are available to any requesting authority and any accrediting agency.



David Batty
Laboratory Director

26 November 2015